



बागमती प्रदेश सरकार

स्वास्थ्य मन्त्रालय

मदन भण्डारी स्वास्थ्य विज्ञान प्रतिष्ठान



हेटौडा अस्पताल

पत्र संख्या: २०८२/०८३

मिति: २०८२-१०-१८

च.नं:

Notice No :- ५३

Regarding the The Procurement of Supply, Delivery and Installation of USG and Echo Machine for MBAHS, Hetauda Hospital Published through PPMO's eGP Portal on 07th January 2026 Following amendment has been made and uploaded under the heading of Addendum

Amendment for Portable Color Doppler Ultrasound Machine

Existing Technical Specifications	Amended Technical Specifications
Frame rate: not less than 350fps.	Frame rate: not less than 2000fps.
Future Upgradable: (Please quote the upgradable package in separate sheet) a) Strain Echo. b) Elastography Software Package. c) Elastography Software Package. d) BIRAD software	Future Upgradable: (Please quote the upgradable package in separate sheet) a) Strain Echo. b) Elastography Software Package. c) Elastography Software Package. d) BIRAD software e) TEE Probe
Accessories, spares and consumables included: 1. Patient circuit neonate- 1 set 2. Factory made Ventilator Trolley – 1 (local trolley not accepted) 3. Test lung (neonatal) – 1 4. Patient circuit hanger – 1 5. O2 pressure hose – 1 6. Air pressure hose – 1 7. Suitable medical air compressor -1 8. Humidifier hanger - 1 9. Servo Control humidifier with complete set - 1 10. Jet type Nebulizer compactable with ventilator set- 1	Removed

फोन नं. ०१७-५२३७९८, ५२०३०५

ईमेल: hospital.hetauda@gmail.com, info@hetaudahospital.gov.np, वेबसाइट: <https://www.hetaudahospital.gov.np>



बागमती प्रदेश सरकार

स्वास्थ्य मन्त्रालय

मदन भण्डारी स्वास्थ्य विज्ञान प्रतिष्ठान



हेटौडा अस्पताल

Amendment for High-End Ultrasound Machine

Existing Technical Specifications	Amended Technical Specifications
System processing channels must be more than 75,00,000.	System processing channels must be more than 85,00,000 .
Frame rates should be at least 2700 frames/sec.	Frame rates should be at least 4700 frames/sec.
Future Upgradable: (Please quote the upgradable package in separate sheet) a) 2D Follicle b) Shearwave Imaging with fat quantification tools. c) Battery Backup	Future Upgradable: (Please quote the upgradable package in separate sheet) a) 2D Follicle b) Shearwave Imaging with fat quantification tools. c) Battery Backup d) Single Crystal 4D convex probe
Accessories, spares and consumables included: 1. Patient circuit neonate- 1 set 2. Factory made Ventilator Trolley – 1 (local trolley not accepted) 3. Test lung (neonatal) – 1 4. Patient circuit hanger – 1 5. O2 pressure hose – 1 6. Air pressure hose – 1 7. Suitable medical air compressor -1 8. Humidifier hanger - 1 9. Servo Control humidifier with complete set - 1 10. Jet type Nebulizer compactable with ventilator set- 1	Removed

Amendment for High-End Color Doppler Echo Machine

Existing Technical Specifications	Amended Technical Specifications
System processing channels must be more than 75,00,000.	System processing channels must be more than 85,00,000 .
Frame rates should be at least 2700 frames/sec.	Frame rates should be at least 4700 frames/sec.

फोन नं. ०१०-५२३७९६, ५२०३०४

ईमेल: hospital.hetauda@gmail.com, info@hetaudahospital.gov.np, वेबसाइट: <https://www.hetaudahospital.gov.np>



बागमती प्रदेश सरकार

स्वास्थ्य मन्त्रालय

मदन भण्डारी स्वास्थ्य विज्ञान प्रतिष्ठान



हेटौडा अस्पताल

Future Upgradable: (Please quote the upgradable package in separate sheet)
a) TEE Probe (Please specify the frequency range)
b) Stress Echo
c) Strain Echo
d) Arterial Analysis

Future Upgradable: (Please quote the upgradable package in separate sheet)
a) TEE Probe (Please specify the frequency range)
b) Stress Echo
c) Strain Echo
d) Arterial Analysis
e) Heart Assist for Adult



Bhim Sagar Guragain
Registrar

फोन नं. ०१६-५२३७९८, ५२०३०५

ईमेल: hospital.hetauda@gmail.com, info@hetaudahospital.gov.np, वेबसाइट: <https://www.hetaudahospital.gov.np>